



SUPER-FAST MINI THERMOMETER



Operating Instructions

INSTRUMENT OPERATION - The instrument is switched on by removing the probe cover or by pressing the Hold or Max/Min button. Insert the probe into the substance being measured. The instrument reading may take a few seconds to stabilise, depending on the nature of the measurement. The instrument will automatically turn off after one hour of inactivity if not placed within the probe cover.

HOLD - Press the Hold button to retain the displayed reading, °C or °F will flash, press the Hold button again to resume taking temperature. **Please note:** Once Hold is selected the unit will not automatically turn off.

MAX/MIN - Press the Max/Min button to display the maximum reading and a 'Max' symbol will flash in the display. Press the Max/Min button again to display the minimum reading and a 'Min' symbol will flash in the display. Press the Max/Min button again to return to normal readings. To reset, simply hold the Max/Min button for three seconds. Please note: Max/Min readings are not saved, so switching off the instrument will clear these readings.

°C/°F - Unscrew the battery cover at the end of the instrument and remove the battery. To select the desired temperature scale, slide the switch to °C or °F.

BATTERY REPLACEMENT - The battery will need to be replaced when 'Bat' is displayed or if the display dims. Replace the battery with a 3 volt lithium coin cell (CR1632). The battery is located under the battery cover at the end of the instrument. Once the end cap has been removed by unscrewing anticlockwise, remove the battery cover using a small crosshead screwdriver. Once the battery is replaced, screw back down the battery cover and then re-tighten the end cap ensuring the gasket remains in place for continued water resistance of the product.

BATTERY WARNING: KEEP OUT OF REACH OF CHILDREN.
If you think batteries might have been swallowed or placed inside any part of the body, seek immediate medical attention.



EMC/RFI - Instrument performance may be affected if operated within a high frequency radio field, such as near a mobile phone, or if subjected to an electrostatic shock.

GUARANTEE - This instrument carries a one-year guarantee against defects in either components or workmanship. During this period, products that prove to be defective will, at the discretion of ETI, be either repaired or replaced without charge. The product guarantee does not cover damage caused by fair wear and tear, abnormal storage conditions, incorrect use, accidental misuse, abuse, neglect, misapplication or modification. Full details of liability are available within ETI's Terms & Conditions of Sale at etilt.com. In line with our policy of continuous development, we reserve the right to amend our product specification without prior notice.



Supplied by
Electronic Temperature Instruments Ltd
Worthing · West Sussex · BN14 8HQ
01903 202151 · sales@etilt.com · etilt.com

