



## **CERTIFICATE OF CONFORMITY**

Order code: **825-882, 884, 886**

### **Wall-Mounted Thermometer & Hygrometer**

Range: T: -40 to 80°C; H: 0 to 100%rh

Resolution: T: 0.1°C/°F; H: 0.1%rh

Accuracy: T:  $\pm 0.5^{\circ}\text{C}$  (-10 to 50°C); H:  $\pm 3\%$ : (30 to 80%rh)

Made in China

### **DECLARATION OF CONFORMITY**

Conforms to RoHS & REACH Legislation

Also conforms to the following regulations; EC Reg 1935/2004, Directive 2011/65/EU & amendment 2015/863, EU Reg 2014/35/EU, EU Reg 2014/30/EU, EMC

We certify that the goods shown on this certificate are manufactured, tested, inspected and stored, to conform in all respects to the stated specifications and in accordance with the Quality System of ETI Limited, accredited to BS EN ISO 9001, as approved and audited by the British Standards Institution.

Where appropriate and required, the product complies with EEC Directives with regard to European Legislation and CE Marking requirements.

All ETI supplied instruments are suitable for use within HACCP plans where the quoted specification for critical control points match the performance of the instrument.

Signed, on behalf of Electronic Temperature Instruments Ltd

A handwritten signature in black ink, appearing to be 'Andy Reid'.

Andy Reid  
Quality Manager  
Valid as of date of purchase.

Riverside House  
Easting Close  
Worthing  
West Sussex  
BN14 8HQ

T +44 (0)1903 202151  
E [sales@etiltld.com](mailto:sales@etiltld.com)  
W [etiltld.com](http://etiltld.com)