



CERTIFICATE OF CONFORMITY

Order code: **293-010,020,050,070**

ThermaData® Lite

Blind Recording Thermometers

Range: -40 to 85°C

Resolution: 0.1°C

Accuracy: 0.5°C

Made in Britain

Conforms to RoHS & REACH Legislation

Also conforms to the following regulations; BS EN13485, Directive 2011/65/EU & amendment 2015/863, EU Reg 2014/35/EU, EU Reg 2014/30/EU.

DECLARATION OF CONFORMITY

We certify that the goods shown on this certificate are manufactured, tested, inspected and stored, to conform in all respects to the stated specifications and in accordance with the Quality System of ETI Limited, accredited to BS EN ISO 9001, as approved and audited by the British Standards Institution.

Where appropriate and required, the product complies with EEC Directives with regard to European Legislation and UKCA/CE Marking requirements. This product conforms to EN 13485, Air, S, D, 1, -40 to 85°C.

All ETI supplied instruments are suitable for use within HACCP plans where the quoted specification for critical control points match the performance of the instrument.

Signed, on behalf of Electronic Temperature Instruments Ltd

A handwritten signature in black ink, appearing to read 'Andy Reid', is positioned above the printed name.

Andy Reid

Quality Manager

Valid as of date of purchase.

Easting Close
Worthing
West Sussex
BN14 8HQ

T +44 (0)1903 202151
E sales@etiltd.com
W etiltd.com