



CERTIFICATE OF CONFORMITY

Order code: 825-283

HumidityPod

Range: T: -25 to 85°C ; H: 0-99% Rh
Resolution: T: 0.1 °C between -19.9 to 85.0 °C; H: 1% Rh 0-99%
Accuracy: T: $\pm 0.5^{\circ}\text{C}$ (-10 to 70°C otherwise $\pm 1^{\circ}\text{C}$)
H: $\pm 3\%$ (30 to 80% otherwise $\pm 5\%$)

DECLARATION OF CONFORMITY

Conforms to RoHS & REACH Legislation

Also conforms to the following regulations; EC Reg 1935/2004, Directive 2011/65/EU & amendment 2015/863, EU Reg 2014/35/EU, EU Reg 2014/30/EU, EMC

We certify that the goods shown on this certificate are manufactured, tested, inspected and stored, to conform in all respects to the stated specifications and in accordance with the Quality System of ETI Limited, accredited to BS EN ISO 9001, as approved and audited by the British Standards Institution.

Where appropriate and required, the product complies with EEC Directives with regard to European Legislation and CE Marking requirements. Including EC Regulation No 1935/2004 on Materials and articles intended to come into contact with food, The probe shaft being manufactured from 316L Stainless Steel is the intended material for contact with food on the product listed above, Migration testing is Not Applicable.

All ETI supplied instruments are suitable for use within HACCP plans where the quoted specification for critical control points match the performance of the instrument.

Signed, on behalf of Electronic Temperature Instruments Ltd

A handwritten signature in blue ink, appearing to read 'Andy Reid'.

Andy Reid
Quality Manager
Valid as of date of purchase.

Riverside House
Easting Close
Worthing
West Sussex
BN14 8HQ

T +44 (0)1903 202151
E sales@etiltd.com
W etiltd.com